

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092018\
 Data File : VV007671.D
 Acq On : 20 Sep 2018 22:47
 Operator : SY/MD
 Sample : J4989-09
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ET0S1

Quant Time: Sep 21 03:48:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Sep 21 03:27:29 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	109310	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	99398	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	53900	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	17587	3.92	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.40%
7) Chloroethane-d5	1.59	69	19201	4.55	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.00%
11) 1,1-Dichloroethene-d2	2.13	63	34711	2.91	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	58.20%#
20) 2-Butanone-d5	3.96	46	52889	46.15	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.30%
24) Chloroform-d	4.41	84	45001	4.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.00%
26) 1,2-Dichloroethane-d4	5.09	65	23562	4.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.00%
32) Benzene-d6	5.10	84	87039	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.40%
36) 1,2-Dichloropropane-d6	6.12	67	25814	4.44	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.80%
41) Toluene-d8	7.36	98	87362	4.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.00%
43) trans-1,3-Dichloropropene-	7.67	79	8935	3.82	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.40%
46) 2-Hexanone-d5	8.14	63	35993	39.51	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	79.02%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	20915	4.34	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	39273	4.41	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092018\
Data File : VV007671.D
Acq On : 20 Sep 2018 22:47
Operator : SY/MD
Sample : J4989-09
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETOS1

Quant Time: Sep 21 03:48:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Sep 21 03:27:29 2018
Response via : Initial Calibration

