

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW092021\
 Data File : VW022225.D
 Acq On : 20 Sep 2021 12:16
 Operator : SY/MD
 Sample : M3787-09
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0CH1

Quant Time: Sep 21 04:42:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Sep 21 04:41:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	317426	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	301264	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	128188	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	108324	49.278	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	98.560%
7) Chloroethane-d5	1.568	69	86815	50.572	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.140%
11) 1,1-Dichloroethene-d2	2.104	63	148739	40.913	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.820%
21) 2-Butanone-d5	3.902	46	153833	102.040	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.040%
24) Chloroform-d	4.352	84	192102	47.927	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.860%
26) 1,2-Dichloroethane-d4	5.037	65	118209	50.069	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.140%
32) Benzene-d6	5.050	84	400714	49.039	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.080%
36) 1,2-Dichloropropane-d6	6.072	67	126507	49.229	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.460%
41) Toluene-d8	7.320	98	351648	45.589	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.180%
43) trans-1,3-Dichloroprop...	7.625	79	52329	42.784	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.560%
47) 2-Hexanone-d5	8.091	63	90995	92.286	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.290%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	147559	44.835	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.660%
66) 1,2-Dichlorobenzene-d4	11.625	152	126574	50.227	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.460%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092021\
Data File : VV022225.D
Acq On : 20 Sep 2021 12:16
Operator : SY/MD
Sample : M3787-09
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0CH1

Quant Time: Sep 21 04:42:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Sep 21 04:41:26 2021
Response via : Initial Calibration

