

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW092021\
 Data File : VW022237.D
 Acq On : 20 Sep 2021 17:02
 Operator : SY/MD
 Sample : M3788-04
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0CK2

Quant Time: Sep 21 04:44:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Sep 21 04:41:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	303150	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	289140	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	125427	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	92190	43.913	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.820%
7) Chloroethane-d5	1.561	69	77471	47.255	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.500%
11) 1,1-Dichloroethene-d2	2.105	63	130646	37.629	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.260%
21) 2-Butanone-d5	3.879	46	134373	93.329	ug/L	-0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.330%
24) Chloroform-d	4.352	84	174945	45.702	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.400%
26) 1,2-Dichloroethane-d4	5.037	65	108380	48.068	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.140%
32) Benzene-d6	5.053	84	362932	46.278	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.560%
36) 1,2-Dichloropropane-d6	6.072	67	116715	47.323	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.640%
41) Toluene-d8	7.320	98	318788	43.061	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.120%
43) trans-1,3-Dichloroprop...	7.625	79	46665	39.753	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.500%
47) 2-Hexanone-d5	8.091	63	81749	86.385	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	86.390%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	132556	41.965	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	83.940%
66) 1,2-Dichlorobenzene-d4	11.628	152	117342	47.589	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.180%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092021\
Data File : VV022237.D
Acq On : 20 Sep 2021 17:02
Operator : SY/MD
Sample : M3788-04
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0CK2

Quant Time: Sep 21 04:44:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Sep 21 04:41:26 2021
Response via : Initial Calibration

