

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW092021\
 Data File : VW022252.D
 Acq On : 20 Sep 2021 23:27
 Operator : SY/MD
 Sample : M3788-14
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0CL7

Quant Time: Sep 21 06:23:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Sep 21 06:08:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	307349	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	289290	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	122132	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	99827	46.902	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.800%
7) Chloroethane-d5	1.564	69	81743	49.179	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.360%
11) 1,1-Dichloroethene-d2	2.105	63	139990	39.769	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.540%
21) 2-Butanone-d5	3.908	46	142729	97.779	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.780%
24) Chloroform-d	4.352	84	179860	46.344	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.680%
26) 1,2-Dichloroethane-d4	5.034	65	111784	48.900	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.800%
32) Benzene-d6	5.053	84	372502	47.473	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.940%
36) 1,2-Dichloropropane-d6	6.072	67	118710	48.107	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.220%
41) Toluene-d8	7.317	98	323504	43.676	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.360%
43) trans-1,3-Dichloroprop...	7.625	79	45970	39.140	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.280%
47) 2-Hexanone-d5	8.095	63	85649	90.459	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.460%
56) 1,1,2,2-Tetrachloroeth...	10.220	84	136578	43.216	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.440%
66) 1,2-Dichlorobenzene-d4	11.625	152	117779	49.055	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.100%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092021\
Data File : VV022252.D
Acq On : 20 Sep 2021 23:27
Operator : SY/MD
Sample : M3788-14
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0CL7

Quant Time: Sep 21 06:23:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Sep 21 06:08:34 2021
Response via : Initial Calibration

