

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092023\
 Data File : VV032148.D
 Acq On : 20 Sep 2023 15:33
 Operator : SY/MD
 Sample : 04472-16ME
 Misc : 7.02g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EZYM5ME

Quant Time: Sep 21 01:40:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 21 01:37:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	304928	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	301052	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	166763	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	96766	50.707	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.420%
7) Chloroethane-d5	1.503	69	25344	15.150	ug/L	-0.03
Spiked Amount	50.000	Range	70 - 130	Recovery	=	30.300%#
11) 1,1-Dichloroethene-d2	2.037	65	38311	42.873	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.740%
21) 2-Butanone-d5	3.799	46	131890	87.873	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	87.870%
24) Chloroform-d	4.249	84	166847	40.000	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.000%
26) 1,2-Dichloroethane-d4	4.941	65	107346	38.780	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	77.560%
32) Benzene-d6	4.957	84	351638	42.805	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.620%
36) 1,2-Dichloropropane-d6	5.989	67	118065	44.653	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.300%
41) Toluene-d8	7.246	98	317692	41.972	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.940%
43) trans-1,3-Dichloroprop...	7.558	79	54072	39.456	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.920%
47) 2-Hexanone-d5	8.040	63	90638	111.843	ug/L	0.01
Spiked Amount	100.000	Range	45 - 130	Recovery	=	111.840%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	158898	46.505	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.020%
66) 1,2-Dichlorobenzene-d4	11.567	152	134497	44.242	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.480%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092023\
Data File : VV032148.D
Acq On : 20 Sep 2023 15:33
Operator : SY/MD
Sample : 04472-16ME
Misc : 7.02g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EZY5ME

Quant Time: Sep 21 01:40:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 21 01:37:00 2023
Response via : Initial Calibration

