

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW092023\
 Data File : VW032151.D
 Acq On : 20 Sep 2023 17:05
 Operator : SY/MD
 Sample : 04385-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Sep 21 01:40:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 21 01:37:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	264038	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	257324	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	132039	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	77755	47.055	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.100%
7) Chloroethane-d5	1.532	69	64591	44.590	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.180%
11) 1,1-Dichloroethene-d2	2.060	65	33093	42.769	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.540%
21) 2-Butanone-d5	3.789	46	112588	86.630	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	86.630%
24) Chloroform-d	4.252	84	138398	38.318	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	76.640%
26) 1,2-Dichloroethane-d4	4.947	65	89334	37.270	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	74.540%
32) Benzene-d6	4.963	84	292544	41.663	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.320%
36) 1,2-Dichloropropane-d6	5.992	67	98016	43.369	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	86.740%
41) Toluene-d8	7.246	98	266775	41.234	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.460%
43) trans-1,3-Dichloroprop...	7.558	79	41639	35.547	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.100%
47) 2-Hexanone-d5	8.027	63	71153	102.720	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.720%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	121775	41.697	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	83.400%
66) 1,2-Dichlorobenzene-d4	11.564	152	101905	42.337	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092023\
Data File : VV032151.D
Acq On : 20 Sep 2023 17:05
Operator : SY/MD
Sample : 04385-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Sep 21 01:40:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 21 01:37:00 2023
Response via : Initial Calibration

