

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092118\
 Data File : VV007699.D
 Acq On : 21 Sep 2018 15:28
 Operator : SY/MD
 Sample : J4989-07
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETOR9

Quant Time: Sep 22 00:22:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Sep 21 03:27:29 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	106859	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	99120	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	52269	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	15625	3.56	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.20%
7) Chloroethane-d5	1.59	69	17443	4.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.60%
11) 1,1-Dichloroethene-d2	2.13	63	30191	2.59	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	51.80%#
20) 2-Butanone-d5	3.97	46	53833	48.05	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.10%
24) Chloroform-d	4.41	84	43544	4.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.20%
26) 1,2-Dichloroethane-d4	5.09	65	22795	4.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.00%
32) Benzene-d6	5.10	84	80718	4.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.20%
36) 1,2-Dichloropropane-d6	6.12	67	24900	4.29	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.80%
41) Toluene-d8	7.36	98	77670	3.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.60%
43) trans-1,3-Dichloropropene-	7.67	79	8672	3.72	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.40%
46) 2-Hexanone-d5	8.14	63	37289	41.05	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.10%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	20297	4.22	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	37064	4.30	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092118\
Data File : VV007699.D
Acq On : 21 Sep 2018 15:28
Operator : SY/MD
Sample : J4989-07
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ET0R9

Quant Time: Sep 22 00:22:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Sep 21 03:27:29 2018
Response via : Initial Calibration

