

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092118\
 Data File : VV007688.D
 Acq On : 21 Sep 2018 10:13
 Operator : SY/MD
 Sample : VV0921WBL01
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK73

Quant Time: Sep 21 23:51:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Sep 20 07:48:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	101547	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	94607	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	50752	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	18084	4.33	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.60%
7) Chloroethane-d5	1.59	69	17992	4.59	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.80%
11) 1,1-Dichloroethene-d2	2.13	63	33979	3.07	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.40%
20) 2-Butanone-d5	3.97	46	54269	50.97	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.94%
24) Chloroform-d	4.41	84	45579	4.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.60%
26) 1,2-Dichloroethane-d4	5.09	65	24051	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.80%
32) Benzene-d6	5.10	84	86777	4.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.60%
36) 1,2-Dichloropropane-d6	6.12	67	26672	4.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.40%
41) Toluene-d8	7.36	98	84321	4.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.20%
43) trans-1,3-Dichloropropene-	7.67	79	8827	3.97	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.40%
46) 2-Hexanone-d5	8.14	63	38494	44.40	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.80%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	21369	4.66	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	39816	4.75	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092118\
Data File : VV007688.D
Acq On : 21 Sep 2018 10:13
Operator : SY/MD
Sample : VV0921WBL01
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK73

Quant Time: Sep 21 23:51:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Sep 20 07:48:21 2018
Response via : Initial Calibration

