

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092120\
 Data File : VV018372.D
 Acq On : 21 Sep 2020 15:38
 Operator : SY/MD
 Sample : L4090-06
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFWX8

Quant Time: Sep 22 05:30:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Sep 22 04:28:24 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	176932	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	176997	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	83705	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	94740	65.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	131.40%
7) Chloroethane-d5	1.58	69	82799	70.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	140.36%#
11) 1,1-Dichloroethene-d2	2.12	63	143523	54.26	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	108.52%
21) 2-Butanone-d5	3.91	46	121607	137.79	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	137.79%#
24) Chloroform-d	4.38	84	183439	73.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	146.64%#
26) 1,2-Dichloroethane-d4	5.06	65	121909	76.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	152.04%#
32) Benzene-d6	5.08	84	367883	72.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	145.20%#
36) 1,2-Dichloropropane-d6	6.09	67	118776	73.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	147.54%#
41) Toluene-d8	7.34	98	325562	70.50	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	141.00%#
43) trans-1,3-Dichloropropene-	7.64	79	57045	69.73	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	139.46%#
47) 2-Hexanone-d5	8.11	63	77824	127.61	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	127.61%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	140495	69.53	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	139.06%#
64) 1,2-Dichlorobenzene-d4	11.65	152	136224	79.24	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	158.48%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092120\
Data File : VV018372.D
Acq On : 21 Sep 2020 15:38
Operator : SY/MD
Sample : L4090-06
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFWX8

Quant Time: Sep 22 05:30:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Sep 22 04:28:24 2020
Response via : Initial Calibration

