

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092120\
 Data File : VV018368.D
 Acq On : 21 Sep 2020 14:08
 Operator : SY/MD
 Sample : L4090-16
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFX01

Quant Time: Sep 22 05:18:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Sep 22 04:28:24 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	227526	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	228395	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	108263	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	91505	49.35	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	98.70%
7) Chloroethane-d5	1.58	69	76342	50.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.62%
11) 1,1-Dichloroethene-d2	2.12	63	133764	39.33	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.66%
21) 2-Butanone-d5	3.91	46	104634	92.19	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.19%
24) Chloroform-d	4.38	84	164717	51.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.40%
26) 1,2-Dichloroethane-d4	5.06	65	109559	53.13	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.26%
32) Benzene-d6	5.08	84	329612	50.41	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.82%
36) 1,2-Dichloropropane-d6	6.09	67	107603	51.79	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.58%
41) Toluene-d8	7.34	98	292222	49.04	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.08%
43) trans-1,3-Dichloropropene-	7.64	79	50020	47.38	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.76%
47) 2-Hexanone-d5	8.11	63	64774	82.31	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	82.31%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	116903	44.83	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.66%
64) 1,2-Dichlorobenzene-d4	11.65	152	122931	55.29	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.58%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092120\
Data File : VV018368.D
Acq On : 21 Sep 2020 14:08
Operator : SY/MD
Sample : L4090-16
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFX01

Quant Time: Sep 22 05:18:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Sep 22 04:28:24 2020
Response via : Initial Calibration

