

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092220\
 Data File : VV018388.D
 Acq On : 22 Sep 2020 12:21
 Operator : SY/MD
 Sample : L4090-12DL 50X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFWY5DL

Quant Time: Sep 23 06:09:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 23 05:49:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	255314	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	252260	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	120758	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	93059	44.72	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.44%
7) Chloroethane-d5	1.58	69	79780	46.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.72%
11) 1,1-Dichloroethene-d2	2.12	63	138279	36.23	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.46%
21) 2-Butanone-d5	3.92	46	120040	94.26	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.26%
24) Chloroform-d	4.38	84	175789	48.69	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.38%
26) 1,2-Dichloroethane-d4	5.06	65	115108	49.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.48%
32) Benzene-d6	5.08	84	346929	48.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.08%
36) 1,2-Dichloropropane-d6	6.09	67	114822	50.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.08%
41) Toluene-d8	7.34	98	302556	45.97	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.94%
43) trans-1,3-Dichloropropene-	7.64	79	53955	46.27	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.54%
47) 2-Hexanone-d5	8.11	63	75899	87.33	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.33%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	133708	46.43	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.86%
64) 1,2-Dichlorobenzene-d4	11.65	152	129632	52.27	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.54%

Target Compounds					Ovalue
34) Trichloroethene	5.94	95	133155	67.772 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092220\
Data File : VV018388.D
Acq On : 22 Sep 2020 12:21
Operator : SY/MD
Sample : L4090-12DL 50X
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFWY5DL

Quant Time: Sep 23 06:09:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 23 05:49:41 2020
Response via : Initial Calibration

