

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092220\
 Data File : VV018394.D
 Acq On : 22 Sep 2020 14:36
 Operator : SY/MD
 Sample : L4090-10
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFWY3

Quant Time: Sep 23 06:38:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 23 05:49:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	238842	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	240661	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	114113	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	92317	47.43	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.86%
7) Chloroethane-d5	1.58	69	78934	49.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.12%
11) 1,1-Dichloroethene-d2	2.12	63	138955	38.92	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.84%
21) 2-Butanone-d5	3.91	46	122196	102.57	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.57%
24) Chloroform-d	4.38	84	180022	53.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.62%
26) 1,2-Dichloroethane-d4	5.06	65	120027	55.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	110.88%
32) Benzene-d6	5.08	84	362504	52.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.24%
36) 1,2-Dichloropropane-d6	6.09	67	117669	53.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	107.50%
41) Toluene-d8	7.34	98	327006	52.08	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.16%
43) trans-1,3-Dichloropropene-	7.64	79	55998	50.34	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	100.68%
47) 2-Hexanone-d5	8.11	63	75726	91.33	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	91.33%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	136652	49.74	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.48%
64) 1,2-Dichlorobenzene-d4	11.65	152	139508	59.53	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	119.06%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092220\
Data File : VV018394.D
Acq On : 22 Sep 2020 14:36
Operator : SY/MD
Sample : L4090-10
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFWY3

Quant Time: Sep 23 06:38:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 23 05:49:41 2020
Response via : Initial Calibration

