

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092220\
 Data File : VV018398.D
 Acq On : 22 Sep 2020 16:29
 Operator : SY/MD
 Sample : L4024-07
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Sep 23 06:53:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 23 05:49:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	253116	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	249050	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	118292	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	84484	40.96	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.92%
7) Chloroethane-d5	1.58	69	73512	43.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.10%
11) 1,1-Dichloroethene-d2	2.12	63	122922	32.49	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.98%
21) 2-Butanone-d5	3.92	46	119178	94.39	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.39%
24) Chloroform-d	4.38	84	161890	45.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.46%
26) 1,2-Dichloroethane-d4	5.06	65	110292	48.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.14%
32) Benzene-d6	5.08	84	319055	44.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.50%
36) 1,2-Dichloropropane-d6	6.09	67	106325	46.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.86%
41) Toluene-d8	7.34	98	276418	42.54	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.08%
43) trans-1,3-Dichloropropene-	7.64	79	47377	41.16	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.32%
47) 2-Hexanone-d5	8.11	63	71682	83.54	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	83.54%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	125156	44.02	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.04%
64) 1,2-Dichlorobenzene-d4	11.65	152	117350	48.30	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092220\
Data File : VV018398.D
Acq On : 22 Sep 2020 16:29
Operator : SY/MD
Sample : L4024-07
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Sep 23 06:53:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 23 05:49:41 2020
Response via : Initial Calibration

