

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092220\
 Data File : VV018386.D
 Acq On : 22 Sep 2020 11:25
 Operator : SY/MD
 Sample : VV0922WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK77

Quant Time: Sep 23 06:01:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 23 05:49:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	275389	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	276744	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	129602	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	94802	42.24	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.48%
7) Chloroethane-d5	1.58	69	80258	43.70	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.40%
11) 1,1-Dichloroethene-d2	2.12	63	140659	34.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.34%
21) 2-Butanone-d5	3.91	46	126168	91.85	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.85%
24) Chloroform-d	4.38	84	179376	46.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.14%
26) 1,2-Dichloroethane-d4	5.06	65	119158	47.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.48%
32) Benzene-d6	5.08	84	357648	45.14	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.28%
36) 1,2-Dichloropropane-d6	6.09	67	115515	45.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.78%
41) Toluene-d8	7.34	98	313924	43.48	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.96%
43) trans-1,3-Dichloropropene-	7.64	79	52754	41.24	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.48%
47) 2-Hexanone-d5	8.11	63	79789	83.68	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	83.68%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	139053	44.01	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.02%
64) 1,2-Dichlorobenzene-d4	11.65	152	132078	49.62	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.24%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092220\
Data File : VV018386.D
Acq On : 22 Sep 2020 11:25
Operator : SY/MD
Sample : VV0922WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK77

Quant Time: Sep 23 06:01:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 23 05:49:41 2020
Response via : Initial Calibration

