

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW092222\
 Data File : VW028077.D
 Acq On : 22 Sep 2022 22:43
 Operator : SY/MD
 Sample : N4652-02DL 5X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EXZ04DL

Quant Time: Sep 23 04:52:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 23 04:38:54 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	175077	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	179869	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	83178	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	94981	5.233	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 104.600%		
7) Chloroethane-d5	1.561	69	78295	5.426	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 108.600%		
11) 1,1-Dichloroethene-d2	2.101	63	135643	4.309	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 86.200%		
20) 2-Butanone-d5	3.870	46	152763	49.853	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	= 99.700%		
24) Chloroform-d	4.343	84	150252	5.188	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 103.800%		
26) 1,2-Dichloroethane-d4	5.024	65	73220	5.313	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 106.200%		
32) Benzene-d6	5.043	84	295383	5.008	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 100.200%		
36) 1,2-Dichloropropane-d6	6.066	67	97416	5.298	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 106.000%		
41) Toluene-d8	7.314	98	251749	4.716	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 94.400%		
43) trans-1,3-Dichloroprop...	7.622	79	30701	4.623	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 92.400%		
46) 2-Hexanone-d5	8.085	63	132571	47.826	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 95.660%		
56) 1,1,2,2-Tetrachloroeth...	10.214	84	62722	4.991	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 99.800%		
66) 1,2-Dichlorobenzene-d4	11.622	152	87798	5.381	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 107.600%		
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.114	96	10311	0.852	ug/L	# 1
16) Methylene chloride	2.500	84	9258	0.685	ug/L	92
18) trans-1,2-Dichloroethene	2.754	96	1034	0.098	ug/L	92
19) 1,1-Dichloroethane	3.182	63	67521	2.922	ug/L	99
22) cis-1,2-Dichloroethene	3.902	96	8764	0.631	ug/L	76
25) Chloroform	4.362	83	3860	0.156	ug/L	94
29) 1,1,1-Trichloroethane	4.600	97	272608	12.335	ug/L	99
34) Trichloroethene	5.918	95	5659	0.455	ug/L	94
47) Tetrachloroethene	7.976	164	8977	0.905	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092222\
Data File : VV028077.D
Acq On : 22 Sep 2022 22:43
Operator : SY/MD
Sample : N4652-02DL 5X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EXZ04DL

Quant Time: Sep 23 04:52:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 23 04:38:54 2022
Response via : Initial Calibration

