

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092318\
 Data File : VV007745.D
 Acq On : 22 Sep 2018 19:58
 Operator : SY/MD
 Sample : J5012-07
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C08M0

Quant Time: Sep 24 04:10:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Sep 24 02:35:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	190272	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	179378	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	94078	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	40245	38.92	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	77.84%
7) Chloroethane-d5	1.59	69	39650	42.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.38%
11) 1,1-Dichloroethene-d2	2.14	63	75824	31.10	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.20%
21) 2-Butanone-d5	3.95	46	58596	85.79	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.79%
24) Chloroform-d	4.41	84	89308	40.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.68%
26) 1,2-Dichloroethane-d4	5.09	65	61701	43.69	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.38%
32) Benzene-d6	5.11	84	179651	43.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.14%
36) 1,2-Dichloropropane-d6	6.13	67	53103	43.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.22%
41) Toluene-d8	7.36	98	181194	42.05	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.10%
43) trans-1,3-Dichloropropene-	7.67	79	21578	35.15	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.30%
47) 2-Hexanone-d5	8.14	63	45295	84.48	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.48%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	78606	41.36	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.72%
64) 1,2-Dichlorobenzene-d4	11.68	152	84262	43.54	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.08%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092318\
Data File : VV007745.D
Acq On : 22 Sep 2018 19:58
Operator : SY/MD
Sample : J5012-07
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C08M0

Quant Time: Sep 24 04:10:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092318WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Sep 24 02:35:10 2018
Response via : Initial Calibration

