

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092318\
 Data File : VV007756.D
 Acq On : 23 Sep 2018 00:31
 Operator : SY/MD
 Sample : J5012-04 250X
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C08K7

Quant Time: Sep 24 04:46:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Sep 24 02:35:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	167760	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	160499	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	84727	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	38060	41.74	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.48%
7) Chloroethane-d5	1.59	69	39263	47.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.78%
11) 1,1-Dichloroethene-d2	2.14	63	70261	32.68	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.36%
21) 2-Butanone-d5	3.95	46	60322	100.17	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.17%
24) Chloroform-d	4.41	84	88199	45.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.36%
26) 1,2-Dichloroethane-d4	5.09	65	60774	48.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.62%
32) Benzene-d6	5.11	84	173674	47.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.16%
36) 1,2-Dichloropropane-d6	6.13	67	52915	48.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.14%
41) Toluene-d8	7.36	98	168702	43.76	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.52%
43) trans-1,3-Dichloropropene-	7.67	79	21073	38.37	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.74%
47) 2-Hexanone-d5	8.14	63	46156	96.21	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.21%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	79073	46.50	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	82377	47.26	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.52%

Target Compounds

					Ovalue
33) Benzene	5.15	78	148373	38.594	ug/L 100
51) Chlorobenzene	8.93	112	357059	108.774	ug/L 99
62) 1,3-Dichlorobenzene	11.23	146	25277	9.409	ug/L 98
63) 1,4-Dichlorobenzene	11.32	146	171541	61.562	ug/L 99
65) 1,2-Dichlorobenzene	11.69	146	207886	76.531	ug/L 100
68) 1,2,4-trichlorobenzene	13.32	180	21675	10.704	ug/L 99
70) 1,2,3-Trichlorobenzene	13.80	180	7666	3.745	ug/L 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092318\
Data File : VV007756.D
Acq On : 23 Sep 2018 00:31
Operator : SY/MD
Sample : J5012-04 250X
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C08K7

Quant Time: Sep 24 04:46:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092318WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Sep 24 02:35:10 2018
Response via : Initial Calibration

