

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092318\
 Data File : VV007779.D
 Acq On : 23 Sep 2018 11:15
 Operator : SY/MD
 Sample : VV0923WBL03
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK64

Quant Time: Sep 24 07:35:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Sep 24 07:19:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	189912	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	184331	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	99065	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	48364	46.86	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.72%
7) Chloroethane-d5	1.59	69	47503	50.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.30%
11) 1,1-Dichloroethene-d2	2.14	63	92406	37.97	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.94%
21) 2-Butanone-d5	3.95	46	64257	94.26	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.26%
24) Chloroform-d	4.41	84	102513	46.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.78%
26) 1,2-Dichloroethane-d4	5.09	65	70145	49.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.54%
32) Benzene-d6	5.10	84	207855	49.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.12%
36) 1,2-Dichloropropane-d6	6.12	67	61557	49.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.40%
41) Toluene-d8	7.36	98	211644	47.80	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.60%
43) trans-1,3-Dichloropropene-	7.67	79	28039	44.45	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.90%
47) 2-Hexanone-d5	8.14	63	45202	82.04	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	82.04%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	88549	45.34	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.68%
64) 1,2-Dichlorobenzene-d4	11.68	152	98600	48.38	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.76%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092318\
Data File : VV007779.D
Acq On : 23 Sep 2018 11:15
Operator : SY/MD
Sample : VV0923WBL03
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 57 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK64

Quant Time: Sep 24 07:35:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092318WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Sep 24 07:19:48 2018
Response via : Initial Calibration

