

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092321\
 Data File : VV022283.D
 Acq On : 23 Sep 2021 14:54
 Operator : SY/MD
 Sample : M3814-19
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GB6X9

Quant Time: Sep 24 03:45:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 24 03:31:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	122758	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	125256	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	56476	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	25502	4.123	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	82.400%	
7) Chloroethane-d5	1.565	69	25354	3.503	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	70.000%	
11) 1,1-Dichloroethene-d2	2.105	63	38159	2.587	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	51.800%#	
20) 2-Butanone-d5	3.880	46	110867	43.860	ug/L	-0.03
Spiked Amount	50.000	Range 40 - 130	Recovery	=	87.720%	
24) Chloroform-d	4.352	84	65775	3.707	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	74.200%	
26) 1,2-Dichloroethane-d4	5.037	65	34104	4.023	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	80.400%	
32) Benzene-d6	5.053	84	127541	3.621	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	72.400%	
36) 1,2-Dichloropropane-d6	6.072	67	42876	3.898	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	78.000%	
41) Toluene-d8	7.320	98	109943	3.457	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	69.200%#	
43) trans-1,3-Dichloroprop...	7.629	79	12330	3.670	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	73.400%	
46) 2-Hexanone-d5	8.088	63	86336	41.333	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	82.660%	
56) 1,1,2,2-Tetrachloroeth...	10.220	84	37681	4.069	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	81.400%	
66) 1,2-Dichlorobenzene-d4	11.628	152	45220	4.266	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	85.400%	
Target Compounds						Qvalue
13) Acetone	2.166	43	20582	14.361	ug/L	100
42) Toluene	7.400	91	4552	0.130	ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092321\
Data File : VV022283.D
Acq On : 23 Sep 2021 14:54
Operator : SY/MD
Sample : M3814-19
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GB6X9

Quant Time: Sep 24 03:45:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 24 03:31:01 2021
Response via : Initial Calibration

