

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW092321\
 Data File : VW022293.D
 Acq On : 23 Sep 2021 18:56
 Operator : SY/MD
 Sample : M3873-21
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW7Y8

Quant Time: Sep 24 03:46:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 24 03:31:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	118695	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	119301	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	52222	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	29095	4.864	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.200%
7) Chloroethane-d5	1.564	69	35011	5.002	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.000%
11) 1,1-Dichloroethene-d2	2.105	63	50000	3.506	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.200%
20) 2-Butanone-d5	3.905	46	129441	52.961	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.920%
24) Chloroform-d	4.349	84	84518	4.927	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.600%
26) 1,2-Dichloroethane-d4	5.037	65	42398	5.173	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.400%
32) Benzene-d6	5.053	84	167201	4.984	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.600%
36) 1,2-Dichloropropane-d6	6.069	67	55126	5.262	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.200%
41) Toluene-d8	7.320	98	139747	4.613	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.200%
43) trans-1,3-Dichloroprop...	7.628	79	14079	4.400	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.000%
46) 2-Hexanone-d5	8.091	63	104334	52.443	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.880%
56) 1,1,2,2-Tetrachloroeth...	10.220	84	43439	4.925	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.600%
66) 1,2-Dichlorobenzene-d4	11.628	152	53582	5.467	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092321\
Data File : VV022293.D
Acq On : 23 Sep 2021 18:56
Operator : SY/MD
Sample : M3873-21
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW7Y8

Quant Time: Sep 24 03:46:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 24 03:31:01 2021
Response via : Initial Calibration

