

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092418\
 Data File : VV007806.D
 Acq On : 24 Sep 2018 11:00
 Operator : SY/MD
 Sample : J5010-08DL 5X
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0T14DL

Quant Time: Sep 25 02:27:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Sep 25 02:11:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	60823	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	60076	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	32478	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	55576	168.12	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	336.24%#
7) Chloroethane-d5	1.59	69	56205	187.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	374.20%#
11) 1,1-Dichloroethene-d2	2.14	63	106683	136.88	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	273.76%#
21) 2-Butanone-d5	3.95	46	75056	343.78	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	343.78%#
24) Chloroform-d	4.41	84	124980	176.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	353.18%#
26) 1,2-Dichloroethane-d4	5.09	65	86490	191.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	383.18%#
32) Benzene-d6	5.10	84	254651	184.41	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	368.82%#
36) 1,2-Dichloropropane-d6	6.12	67	75460	185.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	370.10%#
41) Toluene-d8	7.36	98	257788	178.65	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	357.30%#
43) trans-1,3-Dichloropropene-	7.67	79	34718	168.87	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	337.74%#
47) 2-Hexanone-d5	8.14	63	54347	302.64	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	302.64%#
57) 1,1,2,2-Tetrachloroethane-	10.26	84	106338	167.05	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	334.10%#
64) 1,2-Dichlorobenzene-d4	11.68	152	119229	178.46	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	356.92%#

Target Compounds					Ovalue
9) Trichlorofluoromethane	1.78	101	5433	6.701 ug/L	98
34) Trichloroethene	5.96	95	51489	120.984 ug/L	97
46) Tetrachloroethene	8.02	164	6460	14.831 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092418\
Data File : VV007806.D
Acq On : 24 Sep 2018 11:00
Operator : SY/MD
Sample : J5010-08DL 5X
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0T14DL

Quant Time: Sep 25 02:27:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092318WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Sep 25 02:11:24 2018
Response via : Initial Calibration

