

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092418\
 Data File : VV007828.D
 Acq On : 24 Sep 2018 20:13
 Operator : SY/MD
 Sample : J5043-03MSD
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E5FE2MSD

Quant Time: Sep 25 04:44:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Sep 25 04:04:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	223024	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	216951	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	113558	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	40330	33.27	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	66.54%
7) Chloroethane-d5	1.58	69	44426	40.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	80.66%
11) 1,1-Dichloroethene-d2	2.14	63	117049	40.96	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.92%
21) 2-Butanone-d5	3.95	46	66497	83.06	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	83.06%
24) Chloroform-d	4.41	84	105115	40.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.00%
26) 1,2-Dichloroethane-d4	5.09	65	71655	43.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.58%
32) Benzene-d6	5.11	84	203498	40.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.62%
36) 1,2-Dichloropropane-d6	6.13	67	63140	42.88	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	85.76%
41) Toluene-d8	7.36	98	207351	39.79	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	79.58%#
43) trans-1,3-Dichloropropene-	7.67	79	28767	38.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.50%
47) 2-Hexanone-d5	8.14	63	51250	79.03	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	79.03%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	96782	42.10	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	100493	43.02	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.04%

Target Compounds

					Ovalue
12) 1,1-Dichloroethene	2.15	96	70474	47.642	ug/L 94
33) Benzene	5.15	78	248457	47.811	ug/L 100
34) Trichloroethene	5.96	95	68328	44.458	ug/L 99
42) Toluene	7.44	91	298768	48.502	ug/L 100
51) Chlorobenzene	8.93	112	206985	46.648	ug/L 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092418\
Data File : VV007828.D
Acq On : 24 Sep 2018 20:13
Operator : SY/MD
Sample : J5043-03MSD
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E5FE2MSD

Quant Time: Sep 25 04:44:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092318WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Sep 25 04:04:32 2018
Response via : Initial Calibration

