

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092418\
 Data File : VV007840.D
 Acq On : 25 Sep 2018 01:11
 Operator : SY/MD
 Sample : J5043-15
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E5FF5

Quant Time: Sep 25 05:14:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Sep 25 04:04:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	182144	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	172025	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	82374	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	33449	33.79	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	67.58%
7) Chloroethane-d5	1.59	69	37793	42.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.02%
11) 1,1-Dichloroethene-d2	2.14	63	67356	28.86	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	57.72%#
21) 2-Butanone-d5	3.95	46	56550	86.49	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	86.49%
24) Chloroform-d	4.41	84	84727	39.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	79.96%
26) 1,2-Dichloroethane-d4	5.09	65	60657	44.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.74%
32) Benzene-d6	5.11	84	168640	42.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.30%
36) 1,2-Dichloropropane-d6	6.13	67	52161	44.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.34%
41) Toluene-d8	7.36	98	166722	40.35	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	80.70%
43) trans-1,3-Dichloropropene-	7.67	79	22286	37.86	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.72%
47) 2-Hexanone-d5	8.14	63	40242	78.26	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	78.26%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	79260	43.48	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.96%
64) 1,2-Dichlorobenzene-d4	11.68	152	74078	43.72	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.44%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092418\
Data File : VV007840.D
Acq On : 25 Sep 2018 01:11
Operator : SY/MD
Sample : J5043-15
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E5FF5

Quant Time: Sep 25 05:14:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092318WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Sep 25 04:04:32 2018
Response via : Initial Calibration

