

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092418\
 Data File : VV007844.D
 Acq On : 25 Sep 2018 02:50
 Operator : SY/MD
 Sample : J5043-19
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E5FF9

Quant Time: Sep 25 05:28:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Sep 25 04:04:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	190283	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	187414	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	95251	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	33867	32.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	65.50%
7) Chloroethane-d5	1.59	69	36858	39.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.44%
11) 1,1-Dichloroethene-d2	2.14	63	69229	28.39	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	56.78%#
21) 2-Butanone-d5	3.95	46	49080	71.86	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	71.86%
24) Chloroform-d	4.41	84	83867	37.88	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	75.76%
26) 1,2-Dichloroethane-d4	5.09	65	61064	43.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.48%
32) Benzene-d6	5.11	84	173869	40.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.72%
36) 1,2-Dichloropropane-d6	6.13	67	52434	41.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	82.44%
41) Toluene-d8	7.36	98	174219	38.70	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	77.40%#
43) trans-1,3-Dichloropropene-	7.67	79	23060	35.95	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.90%
47) 2-Hexanone-d5	8.14	63	42531	75.92	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	75.92%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	81140	40.86	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	81.72%
64) 1,2-Dichlorobenzene-d4	11.68	152	83884	42.81	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.62%

Target Compounds

					Ovalue
17) trans-1,2-Dichloroethene	2.80	96	2350	1.651 ug/L	99
20) cis-1,2-Dichloroethene	3.97	96	63347	49.607 ug/L	99
34) Trichloroethene	5.96	95	79100	59.579 ug/L	99
46) Tetrachloroethene	8.02	164	378707	278.703 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092418\
Data File : VV007844.D
Acq On : 25 Sep 2018 02:50
Operator : SY/MD
Sample : J5043-19
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E5FF9

Quant Time: Sep 25 05:28:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092318WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Sep 25 04:04:32 2018
Response via : Initial Calibration

