

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092420\
 Data File : VV018430.D
 Acq On : 24 Sep 2020 15:36
 Operator : SY/MD
 Sample : VIBLK50
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK50

Quant Time: Sep 25 06:12:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 25 04:36:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	245361	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	237170	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	124380	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	109430	54.73	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	109.46%
7) Chloroethane-d5	1.58	69	86843	53.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	106.16%
11) 1,1-Dichloroethene-d2	2.12	63	150858	41.13	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.26%
21) 2-Butanone-d5	3.91	46	127131	103.87	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.87%
24) Chloroform-d	4.38	84	176121	50.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.54%
26) 1,2-Dichloroethane-d4	5.06	65	116718	52.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.96%
32) Benzene-d6	5.08	84	373963	55.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	110.16%
36) 1,2-Dichloropropane-d6	6.09	67	116636	54.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	108.14%
41) Toluene-d8	7.34	98	340775	55.07	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.14%
43) trans-1,3-Dichloropropene-	7.64	79	57697	52.63	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	105.26%
47) 2-Hexanone-d5	8.11	63	77195	94.47	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.47%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	135445	50.02	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.04%
64) 1,2-Dichlorobenzene-d4	11.65	152	142503	55.79	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.58%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092420\
Data File : VV018430.D
Acq On : 24 Sep 2020 15:36
Operator : SY/MD
Sample : VIBLK50
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK50

Quant Time: Sep 25 06:12:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 25 04:36:20 2020
Response via : Initial Calibration

