

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092421\
 Data File : VV022313.D
 Acq On : 24 Sep 2021 11:15
 Operator : SY/MD
 Sample : VV0924WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK211

Quant Time: Sep 25 04:30:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Sep 25 04:30:06 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	140487	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	138677	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	60469	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	27782	3.924	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	78.400%	
7) Chloroethane-d5	1.561	69	34460	4.160	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	83.200%	
11) 1,1-Dichloroethene-d2	2.101	63	52499	3.111	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	62.200%	
20) 2-Butanone-d5	3.912	46	107418	37.132	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	74.260%	
24) Chloroform-d	4.346	84	82549	4.066	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	81.400%	
26) 1,2-Dichloroethane-d4	5.031	65	39235	4.045	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	80.800%	
32) Benzene-d6	5.047	84	162050	4.156	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	83.200%	
36) 1,2-Dichloropropane-d6	6.066	67	53504	4.394	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	87.800%	
41) Toluene-d8	7.317	98	130984	3.720	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	74.400%	
43) trans-1,3-Dichloroprop...	7.625	79	13384	3.598	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	72.000%	
46) 2-Hexanone-d5	8.091	63	82397	35.630	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	71.260%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	37006	3.610	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	72.200%	
66) 1,2-Dichlorobenzene-d4	11.625	152	49938	4.400	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	88.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092421\
Data File : VV022313.D
Acq On : 24 Sep 2021 11:15
Operator : SY/MD
Sample : VV0924WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK211

Quant Time: Sep 25 04:30:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Sep 25 04:30:06 2021
Response via : Initial Calibration

