


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092421\
Data File : VV022314.D
Acq On : 24 Sep 2021 11:49
Operator : SY/MD
Sample : M3878-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW2J7

Manual Integrations
APPROVED

MMDadoda
9/25/2021 10:22:32 AM

Quant Time: Sep 25 04:30:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Sep 25 04:30:06 2021
Response via : Initial Calibration



Quantitation Report (Qedit)




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092421\
 Data File : VV022314.D
 Acq On : 24 Sep 2021 11:49
 Operator : SY/MD
 Sample : M3878-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW2J7

Manual Integrations
 APPROVED

MMDadoda
 9/25/2021 10:22:32 AM

Quant Time: Sep 25 04:30:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Sep 25 04:30:06 2021
 Response via : Initial Calibration



TIC: VV022314.D\data.ms

(13) Acetone (T)

2.191min (-0.016) 2.23 ug/L

response 3251

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	33.10	49.34
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092421\
 Data File : VV022314.D
 Acq On : 24 Sep 2021 11:49
 Operator : SY/MD
 Sample : M3878-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW2J7

Manual Integrations
 APPROVED

MMDadoda
 9/25/2021 10:22:32 AM

Quant Time: Sep 25 04:30:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Sep 25 04:30:06 2021
 Response via : Initial Calibration



TIC: VV022314.D\data.ms

(13) Acetone (T)

2.223min (+ 0.016) 10.58 ug/L m

response 15449

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	33.10	10.38
0.00	0.00	0.00
0.00	0.00	0.00

MD
10/6/21

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\V092421\
 Data File : VV022314.D
 Acq On : 24 Sep 2021 11:49
 Operator : SY/MD
 Sample : M3878-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW2J7

Manual Integrations
 APPROVED

MMDadoda
 9/25/2021 10:22:32 AM

Quant Time: Sep 25 04:30:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Sep 25 04:30:06 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	125113	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	123864	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	53178	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	27218	4.317	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	86.400%	
7) Chloroethane-d5	1.564	69	34452	4.670	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	93.400%	
11) 1,1-Dichloroethene-d2	2.101	63	50010	3.327	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	66.600%	
20) 2-Butanone-d5	3.915	46	104733	40.653	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	81.300%	
24) Chloroform-d	4.346	84	81622	4.514	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.200%	
26) 1,2-Dichloroethane-d4	5.034	65	39421	4.563	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	91.200%	
32) Benzene-d6	5.047	84	162324	4.660	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	93.200%	
36) 1,2-Dichloropropane-d6	6.069	67	52015	4.782	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	95.600%	
41) Toluene-d8	7.316	98	131358	4.176	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	83.600%	
43) trans-1,3-Dichloroprop...	7.628	79	11944	3.595	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	72.000%	
46) 2-Hexanone-d5	8.091	63	82940	40.154	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	80.300%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	35480	3.875	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	77.400%	
66) 1,2-Dichlorobenzene-d4	11.625	152	47831	4.792	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	95.800%	
Target Compounds						
13) Acetone	2.223	43	15449m	10.577	ug/L	Qvalue
42) Toluene	7.403	91	3966	0.115	ug/L	99

> MD
 10/06/21

(#) = qualifier out of range (m) = manual integration (+) = signals summed