

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092424\
 Data File : VV037503.D
 Acq On : 24 Sep 2024 13:42
 Operator : SY/MD
 Sample : VIBLK387
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK387

Quant Time: Sep 25 03:00:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR091024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 25 02:58:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	400340	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	387446	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.182	152	188027	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	85604	3.645	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.800%
7) Chloroethane-d5	1.529	69	81209	3.316	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	66.400%
11) 1,1-Dichloroethene-d2	2.053	65	40097	3.306	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.200%
20) 2-Butanone-d5	3.822	46	261013	37.186	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	74.380%
24) Chloroform-d	4.243	84	206265	3.617	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	72.400%
26) 1,2-Dichloroethane-d4	4.937	65	104547	3.916	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.400%
32) Benzene-d6	4.953	84	406994	3.918	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.400%
36) 1,2-Dichloropropane-d6	5.982	67	130653	4.044	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	80.800%
41) Toluene-d8	7.240	98	357648	3.821	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.400%
43) trans-1,3-Dichloroprop...	7.558	79	41936	3.946	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.000%
46) 2-Hexanone-d5	8.027	63	226383	43.604	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.200%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	108281	4.113	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.200%
66) 1,2-Dichlorobenzene-d4	11.555	152	141198	4.224	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	84.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092424\
Data File : VV037503.D
Acq On : 24 Sep 2024 13:42
Operator : SY/MD
Sample : VIBLK387
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK387

Quant Time: Sep 25 03:00:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR091024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Sep 25 02:58:03 2024
Response via : Initial Calibration

