

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092425\
 Data File : VV039147.D
 Acq On : 24 Sep 2025 12:08
 Operator : SY/MD
 Sample : VV0924MBL01
 Misc : 5.00g/10.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VV0924MBL01

Quant Time: Sep 25 04:43:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	627500	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	1356895	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	1272637	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	538338	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.941	65	578821	63.734	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	127.460%#
35) Dibromofluoromethane	4.500	113	482503	44.232	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	88.460%
50) Toluene-d8	7.236	98	1506826	47.033	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	94.060%
62) 4-Bromofluorobenzene	10.001	95	490201	37.167	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	74.340%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092425\
Data File : VV039147.D
Acq On : 24 Sep 2025 12:08
Operator : SY/MD
Sample : VV0924MBL01
Misc : 5.00g/10.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VV0924MBL01

Quant Time: Sep 25 04:43:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

