

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092425\
 Data File : VV039151.D
 Acq On : 24 Sep 2025 14:46
 Operator : SY/MD
 Sample : Q3173-02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 Field Blank

Quant Time: Sep 25 04:46:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	552493	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	1029152	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	962323	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.172	152	434712	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.941	65	462839	57.882	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	115.760%
35) Dibromofluoromethane	4.500	113	418508	50.584	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	101.160%
50) Toluene-d8	7.236	98	1060720	43.652	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	87.300%#
62) 4-Bromofluorobenzene	10.001	95	446507	44.635	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	89.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092425\
Data File : VV039151.D
Acq On : 24 Sep 2025 14:46
Operator : SY/MD
Sample : Q3173-02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
Field Blank

Quant Time: Sep 25 04:46:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

