

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092425\
 Data File : VV039163.D
 Acq On : 24 Sep 2025 19:15
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK

Quant Time: Sep 25 08:22:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	536805	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	968588	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	977586	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.172	152	475135	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	444222	57.178	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	114.360%
35) Dibromofluoromethane	4.503	113	391937	50.334	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	100.660%
50) Toluene-d8	7.236	98	1073286	46.931	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	93.860%
62) 4-Bromofluorobenzene	10.001	95	411373	43.694	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	87.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092425\
Data File : VV039163.D
Acq On : 24 Sep 2025 19:15
Operator : SY/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK

Quant Time: Sep 25 08:22:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

