

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092518\
 Data File : VV007863.D
 Acq On : 25 Sep 2018 19:16
 Operator : SY/MD
 Sample : J5043-21
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Sep 26 13:05:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092518WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 26 04:58:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	192325	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	184089	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	93701	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	56032	46.48	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	92.96%
7) Chloroethane-d5	1.59	69	52159	50.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.20%
11) 1,1-Dichloroethene-d2	2.14	63	95094	37.02	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.04%
21) 2-Butanone-d5	3.96	46	39228	48.28	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	48.28%
24) Chloroform-d	4.41	84	105740	42.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.46%
26) 1,2-Dichloroethane-d4	5.09	65	88672	49.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.32%
32) Benzene-d6	5.11	84	245422	49.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.56%
36) 1,2-Dichloropropane-d6	6.13	67	79459	51.47	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.94%
41) Toluene-d8	7.36	98	240836	48.46	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.92%
43) trans-1,3-Dichloropropene-	7.67	79	33842	46.04	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.08%
47) 2-Hexanone-d5	8.14	63	60841	93.32	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.32%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	106713	46.83	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.66%
64) 1,2-Dichlorobenzene-d4	11.68	152	103646	49.76	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.52%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092518\
Data File : VV007863.D
Acq On : 25 Sep 2018 19:16
Operator : SY/MD
Sample : J5043-21
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Sep 26 13:05:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092518WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 26 04:58:06 2018
Response via : Initial Calibration

