

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092518\
 Data File : VV007865.D
 Acq On : 25 Sep 2018 20:05
 Operator : SY/MD
 Sample : J5045-15
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Sep 26 13:08:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092518WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 26 04:58:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	186645	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	182248	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	91382	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	55034	47.04	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.08%
7) Chloroethane-d5	1.59	69	51085	50.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.12%
11) 1,1-Dichloroethene-d2	2.14	63	93993	37.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.40%
21) 2-Butanone-d5	3.95	46	68910	87.39	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	87.39%
24) Chloroform-d	4.41	84	108660	44.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.42%
26) 1,2-Dichloroethane-d4	5.09	65	89720	51.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.54%
32) Benzene-d6	5.11	84	245839	50.37	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.74%
36) 1,2-Dichloropropane-d6	6.13	67	77808	50.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.82%
41) Toluene-d8	7.36	98	239496	48.68	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.36%
43) trans-1,3-Dichloropropene-	7.67	79	33450	45.97	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.94%
47) 2-Hexanone-d5	8.14	63	59940	92.86	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.86%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	107290	47.56	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.12%
64) 1,2-Dichlorobenzene-d4	11.68	152	104068	51.23	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.46%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092518\
Data File : VV007865.D
Acq On : 25 Sep 2018 20:05
Operator : SY/MD
Sample : J5045-15
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Sep 26 13:08:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092518WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 26 04:58:06 2018
Response via : Initial Calibration

