

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092518\
 Data File : VV007867.D
 Acq On : 25 Sep 2018 20:55
 Operator : SY/MD
 Sample : J5014-11
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Sep 26 13:09:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092518WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 26 04:58:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	179529	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	178349	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	90488	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	53207	47.28	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.56%
7) Chloroethane-d5	1.59	69	48278	49.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.36%
11) 1,1-Dichloroethene-d2	2.14	63	90848	37.89	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.78%
21) 2-Butanone-d5	3.95	46	77670	102.40	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.40%
24) Chloroform-d	4.41	84	105294	45.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.08%
26) 1,2-Dichloroethane-d4	5.09	65	87192	52.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.62%
32) Benzene-d6	5.11	84	232327	48.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.28%
36) 1,2-Dichloropropane-d6	6.13	67	76021	50.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.64%
41) Toluene-d8	7.36	98	228485	47.46	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.92%
43) trans-1,3-Dichloropropene-	7.67	79	31858	44.74	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.48%
47) 2-Hexanone-d5	8.14	63	61986	98.13	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.13%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	107825	48.84	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.68%
64) 1,2-Dichlorobenzene-d4	11.68	152	100578	50.00	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.00%

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092518\
 Data File : VV007867.D
 Acq On : 25 Sep 2018 20:55
 Operator : SY/MD
 Sample : J5014-11
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampled :
 VHBLK01

Quant Time: Sep 26 13:09:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092518WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 26 04:58:06 2018
 Response via : Initial Calibration

