

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092518\
 Data File : VV007873.D
 Acq On : 25 Sep 2018 23:24
 Operator : SY/MD
 Sample : J5036-04 100X
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C08W8

Quant Time: Sep 26 13:12:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092518WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 26 04:58:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	190208	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	183736	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	91629	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	50533	42.38	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.76%
7) Chloroethane-d5	1.59	69	48225	46.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.68%
11) 1,1-Dichloroethene-d2	2.14	63	88115	34.68	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.36%
21) 2-Butanone-d5	3.95	46	85774	106.74	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.74%
24) Chloroform-d	4.41	84	106973	43.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.38%
26) 1,2-Dichloroethane-d4	5.09	65	85969	48.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.36%
32) Benzene-d6	5.11	84	231115	46.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.94%
36) 1,2-Dichloropropane-d6	6.13	67	74788	48.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.06%
41) Toluene-d8	7.36	98	221496	44.66	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.32%
43) trans-1,3-Dichloropropene-	7.67	79	31164	42.48	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.96%
47) 2-Hexanone-d5	8.14	63	58088	89.26	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.26%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	104101	45.77	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.54%
64) 1,2-Dichlorobenzene-d4	11.68	152	97047	47.64	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.28%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092518\
Data File : VV007873.D
Acq On : 25 Sep 2018 23:24
Operator : SY/MD
Sample : J5036-04 100X
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C08W8

Quant Time: Sep 26 13:12:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092518WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 26 04:58:06 2018
Response via : Initial Calibration

