

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092518\
 Data File : VV007872.D
 Acq On : 25 Sep 2018 22:59
 Operator : SY/MD
 Sample : J5036-03 100X
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C08W7

Quant Time: Sep 26 13:11:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092518WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 26 04:58:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	191305	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	186045	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	91926	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	52715	43.96	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.92%
7) Chloroethane-d5	1.59	69	49930	48.21	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.42%
11) 1,1-Dichloroethene-d2	2.14	63	91324	35.74	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.48%
21) 2-Butanone-d5	3.95	46	77731	96.17	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.17%
24) Chloroform-d	4.41	84	107751	43.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.52%
26) 1,2-Dichloroethane-d4	5.09	65	87836	49.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.90%
32) Benzene-d6	5.11	84	242682	48.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.42%
36) 1,2-Dichloropropane-d6	6.13	67	76851	49.25	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.50%
41) Toluene-d8	7.36	98	233786	46.55	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.10%
43) trans-1,3-Dichloropropene-	7.67	79	32799	44.15	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.30%
47) 2-Hexanone-d5	8.14	63	63487	96.35	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.35%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	106839	46.39	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.78%
64) 1,2-Dichlorobenzene-d4	11.68	152	100886	49.37	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.74%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092518\
Data File : VV007872.D
Acq On : 25 Sep 2018 22:59
Operator : SY/MD
Sample : J5036-03 100X
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C08W7

Quant Time: Sep 26 13:11:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092518WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 26 04:58:06 2018
Response via : Initial Calibration

