

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092520\
 Data File : VV018443.D
 Acq On : 25 Sep 2020 10:16
 Operator : SY/MD
 Sample : VV0925WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK81

Quant Time: Sep 25 23:51:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 25 23:36:18 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	384495	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	369093	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	189544	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	131487	41.96	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.92%
7) Chloroethane-d5	1.58	69	106772	41.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.28%
11) 1,1-Dichloroethene-d2	2.12	63	182851	31.81	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.62%
21) 2-Butanone-d5	3.92	46	156075	81.38	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	81.38%
24) Chloroform-d	4.38	84	219837	40.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.88%
26) 1,2-Dichloroethane-d4	5.06	65	144377	41.43	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.86%
32) Benzene-d6	5.08	84	459543	43.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.98%
36) 1,2-Dichloropropane-d6	6.09	67	145034	43.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	86.40%
41) Toluene-d8	7.34	98	417953	43.40	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.80%
43) trans-1,3-Dichloropropene-	7.64	79	71432	41.87	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.74%
47) 2-Hexanone-d5	8.11	63	95017	74.72	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	74.72%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	170288	40.41	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	80.82%
64) 1,2-Dichlorobenzene-d4	11.65	152	176815	45.42	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.84%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092520\
Data File : VV018443.D
Acq On : 25 Sep 2020 10:16
Operator : SY/MD
Sample : VV0925WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK81

Quant Time: Sep 25 23:51:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 25 23:36:18 2020
Response via : Initial Calibration

