

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092520\
 Data File : VV018469.D
 Acq On : 25 Sep 2020 20:25
 Operator : SY/MD
 Sample : L4149-04
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG3Q3

Quant Time: Sep 26 03:57:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 26 02:48:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	376549	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	360016	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	180506	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	132125	43.06	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.12%
7) Chloroethane-d5	1.58	69	110326	43.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.88%
11) 1,1-Dichloroethene-d2	2.12	63	184639	32.80	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.60%
21) 2-Butanone-d5	3.92	46	171577	91.35	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.35%
24) Chloroform-d	4.38	84	228995	43.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.02%
26) 1,2-Dichloroethane-d4	5.06	65	150123	43.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.98%
32) Benzene-d6	5.08	84	488588	47.41	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.82%
36) 1,2-Dichloropropane-d6	6.09	67	155684	47.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.08%
41) Toluene-d8	7.34	98	440269	46.87	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.74%
43) trans-1,3-Dichloropropene-	7.64	79	72930	43.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.66%
47) 2-Hexanone-d5	8.11	63	101213	81.60	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.60%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	174799	42.53	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.06%
64) 1,2-Dichlorobenzene-d4	11.65	152	182538	49.24	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.48%

Target Compounds					Ovalue
13) Acetone	2.20	43	6788	4.513 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092520\
Data File : VV018469.D
Acq On : 25 Sep 2020 20:25
Operator : SY/MD
Sample : L4149-04
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG3Q3

Quant Time: Sep 26 03:57:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 26 02:48:35 2020
Response via : Initial Calibration

