

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092520\
 Data File : VV018475.D
 Acq On : 25 Sep 2020 22:40
 Operator : SY/MD
 Sample : L4117-08
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0AA9

Quant Time: Sep 26 04:19:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 26 02:48:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	360841	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	346964	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	172639	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	122234	41.57	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.14%
7) Chloroethane-d5	1.58	69	103786	43.13	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.26%
11) 1,1-Dichloroethene-d2	2.12	63	173924	32.24	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.48%
21) 2-Butanone-d5	3.92	46	161256	89.59	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.59%
24) Chloroform-d	4.38	84	220517	43.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.44%
26) 1,2-Dichloroethane-d4	5.06	65	144142	44.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.14%
32) Benzene-d6	5.08	84	460300	46.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.68%
36) 1,2-Dichloropropane-d6	6.09	67	146970	46.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.14%
41) Toluene-d8	7.34	98	410241	45.32	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.64%
43) trans-1,3-Dichloropropene-	7.64	79	67886	42.33	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.66%
47) 2-Hexanone-d5	8.11	63	99313	83.08	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	83.08%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	176680	44.60	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.20%
64) 1,2-Dichlorobenzene-d4	11.65	152	170473	48.08	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.16%

Target Compounds

					Ovalue
17) trans-1,2-Dichloroethene	2.79	96	2326	0.953 ug/L	83
20) cis-1,2-Dichloroethene	3.94	96	68927	26.407 ug/L	93
34) Trichloroethene	5.94	95	487043	180.229 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092520\
Data File : VV018475.D
Acq On : 25 Sep 2020 22:40
Operator : SY/MD
Sample : L4117-08
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0AA9

Quant Time: Sep 26 04:19:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 26 02:48:35 2020
Response via : Initial Calibration

