

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092520\
 Data File : VV018479.D
 Acq On : 26 Sep 2020 00:11
 Operator : SY/MD
 Sample : L4117-12
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0AD2

Quant Time: Sep 26 04:37:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 26 02:48:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	354743	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	342890	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	170020	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	120551	41.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.40%
7) Chloroethane-d5	1.58	69	102898	43.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.00%
11) 1,1-Dichloroethene-d2	2.12	63	170273	32.11	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.22%
21) 2-Butanone-d5	3.92	46	160143	90.50	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.50%
24) Chloroform-d	4.38	84	218790	43.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.24%
26) 1,2-Dichloroethane-d4	5.06	65	141554	44.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.04%
32) Benzene-d6	5.08	84	457051	46.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.12%
36) 1,2-Dichloropropane-d6	6.09	67	144510	46.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.66%
41) Toluene-d8	7.34	98	408648	45.68	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.36%
43) trans-1,3-Dichloropropene-	7.64	79	68823	43.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.84%
47) 2-Hexanone-d5	8.11	63	97949	82.91	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	82.91%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	174852	44.67	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.34%
64) 1,2-Dichlorobenzene-d4	11.65	152	170646	48.87	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.74%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092520\
Data File : VV018479.D
Acq On : 26 Sep 2020 00:11
Operator : SY/MD
Sample : L4117-12
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0AD2

Quant Time: Sep 26 04:37:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 26 02:48:35 2020
Response via : Initial Calibration

