

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092524\
 Data File : VV037530.D
 Acq On : 25 Sep 2024 17:22
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Sep 26 04:14:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 26 04:08:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	629470	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	622684	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	290058	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	195875	34.017	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	68.040%	
7) Chloroethane-d5	1.532	69	171089	38.445	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	76.880%	
11) 1,1-Dichloroethene-d2	2.060	65	90042	37.205	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	74.400%	
21) 2-Butanone-d5	3.796	46	178548	80.072	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	80.070%	
24) Chloroform-d	4.243	84	400848	44.479	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	88.960%	
26) 1,2-Dichloroethane-d4	4.937	65	247700	43.646	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	87.300%	
32) Benzene-d6	4.953	84	789521	41.979	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	83.960%	
36) 1,2-Dichloropropane-d6	5.982	67	259252	44.010	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	88.020%	
41) Toluene-d8	7.236	98	681126	40.128	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	80.260%	
43) trans-1,3-Dichloroprop...	7.551	79	110808	40.659	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	81.320%	
47) 2-Hexanone-d5	8.024	63	117458	73.060	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	73.060%	
56) 1,1,2,2-Tetrachloroeth...	10.146	84	310341	43.439	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	86.880%	
66) 1,2-Dichlorobenzene-d4	11.554	152	273789	47.130	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	94.260%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092524\
Data File : VV037530.D
Acq On : 25 Sep 2024 17:22
Operator : SY/MD
Sample : VIBLK
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Sep 26 04:14:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 26 04:08:51 2024
Response via : Initial Calibration

