

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092524\
 Data File : VV037513.D
 Acq On : 25 Sep 2024 11:06
 Operator : SY/MD
 Sample : P4109-15ME
 Misc : 4.88G/5.0mL/100uL/5mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A4ET7ME

Quant Time: Sep 26 04:10:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 26 04:08:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	476399	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	464481	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.182	152	235420	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	212136	48.678	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	97.360%
7) Chloroethane-d5	1.507	69	97990	29.094	ug/L	-0.03
Spiked Amount	50.000	Range	70 - 130	Recovery	=	58.180%#
11) 1,1-Dichloroethene-d2	2.044	65	95822	52.315	ug/L	-0.01
Spiked Amount	50.000	Range	60 - 125	Recovery	=	104.620%
21) 2-Butanone-d5	3.802	46	177310	105.067	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.070%
24) Chloroform-d	4.240	84	429835	63.021	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	126.040%#
26) 1,2-Dichloroethane-d4	4.934	65	261524	60.888	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	121.780%
32) Benzene-d6	4.947	84	865582	61.699	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	123.400%
36) 1,2-Dichloropropane-d6	5.982	67	278383	63.354	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	126.700%#
41) Toluene-d8	7.236	98	768874	60.726	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	121.460%#
43) trans-1,3-Dichloroprop...	7.551	79	119753	58.907	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	117.820%
47) 2-Hexanone-d5	8.034	63	135659	113.121	ug/L	0.01
Spiked Amount	100.000	Range	45 - 130	Recovery	=	113.120%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	324955	60.976	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	121.960%#
66) 1,2-Dichlorobenzene-d4	11.555	152	313537	66.498	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	133.000%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092524\
Data File : VV037513.D
Acq On : 25 Sep 2024 11:06
Operator : SY/MD
Sample : P4109-15ME
Misc : 4.88G/5.0mL/100uL/5mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A4ET7ME

Quant Time: Sep 26 04:10:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 26 04:08:51 2024
Response via : Initial Calibration

