

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092524\
 Data File : VV037517.D
 Acq On : 25 Sep 2024 12:42
 Operator : SY/MD
 Sample : VIBLK388
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK388

Quant Time: Sep 26 04:12:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 26 04:08:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	727165	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	696268	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	336632	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	224315	33.722	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	67.440%		
7) Chloroethane-d5	1.532	69	194823	37.896	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	75.800%		
11) 1,1-Dichloroethene-d2	2.056	65	102528	36.672	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	73.340%		
21) 2-Butanone-d5	3.789	46	193648	75.176	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	75.180%		
24) Chloroform-d	4.240	84	445900	42.831	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	85.660%		
26) 1,2-Dichloroethane-d4	4.934	65	272286	41.532	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	83.060%		
32) Benzene-d6	4.950	84	906823	43.120	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	86.240%		
36) 1,2-Dichloropropane-d6	5.982	67	288082	43.736	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	87.480%		
41) Toluene-d8	7.236	98	778264	41.005	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	82.000%		
43) trans-1,3-Dichloroprop...	7.548	79	123403	40.495	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	80.980%		
47) 2-Hexanone-d5	8.021	63	123497	68.698	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	68.700%		
56) 1,1,2,2-Tetrachloroeth...	10.146	84	316495	39.618	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	79.240%		
66) 1,2-Dichlorobenzene-d4	11.554	152	306940	45.526	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	91.060%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092524\
Data File : VV037517.D
Acq On : 25 Sep 2024 12:42
Operator : SY/MD
Sample : VIBLK388
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK388

Quant Time: Sep 26 04:12:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 26 04:08:51 2024
Response via : Initial Calibration

