

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092524\
 Data File : VV037514.D
 Acq On : 25 Sep 2024 11:28
 Operator : SY/MD
 Sample : P4109-18ME
 Misc : 4.86G/5.0mL/100uL/5mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A4EW0ME

Quant Time: Sep 26 04:10:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 26 04:08:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	730772	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	701710	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	353005	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	299057	44.737	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.480%
7) Chloroethane-d5	1.510	69	163571	31.660	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	63.320%#
11) 1,1-Dichloroethene-d2	2.047	65	138423	49.267	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.540%
21) 2-Butanone-d5	3.799	46	220235	85.076	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.080%
24) Chloroform-d	4.240	84	597696	57.129	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	114.260%
26) 1,2-Dichloroethane-d4	4.934	65	371174	56.336	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	112.680%
32) Benzene-d6	4.947	84	1213838	57.272	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	114.540%
36) 1,2-Dichloropropane-d6	5.982	67	389823	58.723	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	117.440%
41) Toluene-d8	7.236	98	1078666	56.392	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	112.780%
43) trans-1,3-Dichloroprop...	7.548	79	178167	58.012	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	116.020%
47) 2-Hexanone-d5	8.027	63	201710	111.336	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	111.340%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	447200	55.545	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	111.100%
66) 1,2-Dichlorobenzene-d4	11.554	152	432145	61.124	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	122.240%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092524\
Data File : VV037514.D
Acq On : 25 Sep 2024 11:28
Operator : SY/MD
Sample : P4109-18ME
Misc : 4.86G/5.0mL/100uL/5mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A4EW0ME

Quant Time: Sep 26 04:10:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 26 04:08:51 2024
Response via : Initial Calibration

