

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092524\
 Data File : VV037524.D
 Acq On : 25 Sep 2024 15:13
 Operator : SY/MD
 Sample : P4109-10
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Sep 26 04:13:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 26 04:08:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	611408	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	602181	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	280331	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	204238	36.517	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	73.040%
7) Chloroethane-d5	1.532	69	179011	41.413	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.820%
11) 1,1-Dichloroethene-d2	2.060	65	95709	40.715	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.420%
21) 2-Butanone-d5	3.793	46	190473	87.944	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	87.940%
24) Chloroform-d	4.243	84	420786	48.071	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.140%
26) 1,2-Dichloroethane-d4	4.937	65	263531	47.807	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.620%
32) Benzene-d6	4.953	84	845884	46.507	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.020%
36) 1,2-Dichloropropane-d6	5.982	67	273366	47.986	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.980%
41) Toluene-d8	7.240	98	741831	45.192	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.380%
43) trans-1,3-Dichloroprop...	7.551	79	116475	44.193	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.380%
47) 2-Hexanone-d5	8.024	63	121285	78.009	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	78.010%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	311895	45.143	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.280%
66) 1,2-Dichlorobenzene-d4	11.554	152	284134	50.607	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.220%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092524\
Data File : VV037524.D
Acq On : 25 Sep 2024 15:13
Operator : SY/MD
Sample : P4109-10
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Sep 26 04:13:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 26 04:08:51 2024
Response via : Initial Calibration

