

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092525\
 Data File : VV039169.D
 Acq On : 25 Sep 2025 10:31
 Operator : SY/MD
 Sample : VV0925MBL01
 Misc : 5.0g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VV0925MBL01

Quant Time: Sep 26 01:20:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.597	168	536470	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	954712	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	941473	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.172	152	452663	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.941	65	437769	56.382	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	112.760%
35) Dibromofluoromethane	4.500	113	393702	51.296	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	102.600%
50) Toluene-d8	7.236	98	1070675	47.497	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	95.000%
62) 4-Bromofluorobenzene	10.002	95	404772	43.618	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	87.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092525\
Data File : VV039169.D
Acq On : 25 Sep 2025 10:31
Operator : SY/MD
Sample : VV0925MBL01
Misc : 5.0g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VV0925MBL01

Quant Time: Sep 26 01:20:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

