

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092525\
 Data File : VV039178.D
 Acq On : 25 Sep 2025 14:03
 Operator : SY/MD
 Sample : Q3168-06
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 TOTES COMP#2

Quant Time: Sep 26 01:27:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.603	168	372931	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	727936	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	619945	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	297418	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	410371	76.031	ug/l	0.00
Spiked Amount 50.000	Range 81 - 118		Recovery = 152.060%#			
35) Dibromofluoromethane	4.503	113	350931	59.968	ug/l	0.00
Spiked Amount 50.000	Range 80 - 119		Recovery = 119.940%#			
50) Toluene-d8	7.239	98	690767	40.190	ug/l	0.00
Spiked Amount 50.000	Range 89 - 112		Recovery = 80.380%#			
62) 4-Bromofluorobenzene	10.001	95	287774	40.671	ug/l	0.00
Spiked Amount 50.000	Range 85 - 114		Recovery = 81.340%#			
Target Compounds						Qvalue
11) Tert butyl alcohol	2.571	59	25136	22.034	ug/l	96
16) Acetone	2.121	43	290099	80.374	ug/l	98
17) Carbon Disulfide	2.256	76	6117	0.356	ug/l	95
25) 2-Butanone	3.899	43	64291	14.875	ug/l	100
51) 4-Methyl-2-Pentanone	7.165	43	17594	1.827	ug/l	88
52) Toluene	7.329	92	16560	0.991	ug/l	96
59) 2-Hexanone	8.082	43	56115	9.477	ug/l	96
67) Ethyl Benzene	8.950	91	14526	0.556	ug/l	97
68) m/p-Xylenes	9.072	106	13384	1.327	ug/l	94
69) o-Xylene	9.477	106	7971	0.892	ug/l	93
84) 1,2,4-Trimethylbenzene	10.844	105	20053	1.152	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092525\
Data File : VV039178.D
Acq On : 25 Sep 2025 14:03
Operator : SY/MD
Sample : Q3168-06
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
TOTES COMP#2

Quant Time: Sep 26 01:27:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

