

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
 Data File : VV039188.D  
 Acq On : 25 Sep 2025 17:47  
 Operator : SY/MD  
 Sample : Q3195-24  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 EB-02-092225

Quant Time: Sep 26 01:35:14 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 4.603  | 168  | 315041   | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 5.535  | 114  | 553359   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 8.776  | 117  | 549470   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 11.175 | 152  | 261018   | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |          |
|-----------------------------|--------|-------|----------|----------|------|----------|
| System Monitoring Compounds |        |       |          |          |      |          |
| 33) 1,2-Dichloroethane-d4   | 4.944  | 65    | 252908   | 55.467   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 81 - 118 | Recovery | =    | 110.940% |
| 35) Dibromofluoromethane    | 4.503  | 113   | 229182   | 51.518   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 80 - 119 | Recovery | =    | 103.040% |
| 50) Toluene-d8              | 7.240  | 98    | 600984   | 45.998   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 89 - 112 | Recovery | =    | 92.000%  |
| 62) 4-Bromofluorobenzene    | 10.005 | 95    | 234432   | 43.585   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 85 - 114 | Recovery | =    | 87.180%  |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
Data File : VV039188.D  
Acq On : 25 Sep 2025 17:47  
Operator : SY/MD  
Sample : Q3195-24  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 22 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
EB-02-092225

Quant Time: Sep 26 01:35:14 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration

