

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092525\
 Data File : VV039189.D
 Acq On : 25 Sep 2025 18:10
 Operator : SY/MD
 Sample : Q3195-25
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 FB-01-091925

Quant Time: Sep 26 01:35:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.603	168	562300	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	1002078	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	977857	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	478832	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.941	65	439118	53.958	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	107.920%
35) Dibromofluoromethane	4.500	113	402765	49.996	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	100.000%
50) Toluene-d8	7.240	98	1091605	46.137	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	92.280%
62) 4-Bromofluorobenzene	10.001	95	438151	44.983	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	89.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092525\
Data File : VV039189.D
Acq On : 25 Sep 2025 18:10
Operator : SY/MD
Sample : Q3195-25
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
FB-01-091925

Quant Time: Sep 26 01:35:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

