

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092525\
 Data File : VV039191.D
 Acq On : 25 Sep 2025 18:55
 Operator : SY/MD
 Sample : Q3195-27
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 TB-01-091525

Quant Time: Sep 26 01:36:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.603	168	466916	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	828915	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	809498	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	384637	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	368405	54.517	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	109.040%
35) Dibromofluoromethane	4.503	113	338536	50.802	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	101.600%
50) Toluene-d8	7.239	98	895565	45.758	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	91.520%
62) 4-Bromofluorobenzene	10.001	95	354206	43.962	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	87.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092525\
Data File : VV039191.D
Acq On : 25 Sep 2025 18:55
Operator : SY/MD
Sample : Q3195-27
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
TB-01-091525

Quant Time: Sep 26 01:36:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

